15

AGENT MONITORING PLAN
GB ROCKETS, TONS, AND 105 mm PROJECTILES, AND HD MONITORING FOR LAB P&A

STA #	ACAMS TAG #	DAAMS TAG #	UNIT LOC FLOOR PLAN # DRAWING #	AREA MONITORED FLOOR PLAN #	POWER TYPE SPS/UPS	SAMPLE POINT HAZ CAT	MON-ITORING LEVEL	ALARM LEVEL	SAMPLE LINE LENGTH 10%	DAAMS MODE	COMMENTS
ECR 311	219	128	OBS COR 09-204 01-ID-28-001-9-B4	ECR DUCT "A" RM 03-211	UPS	A	ECL IDLH WITH LVS	400	70'	HIST SYST	NO ALARM HORN
ECR 312	105	120	OBS COR 09-204 01-ID-28-001-9-B4	ECR DUCT "B" RM 03-212	UPS	A	ECL IDLH WITH LVS	400	70'	HIST SYST	NO ALARM HORN
DUC 313	180	N/A	UPA 02-214 01-ID-28-001-9-B2	90" EXH DUCT	UPS	NA	ECL IDLH WITH LVS	400	200'	N/A	MONITORED TO CALCULATE FILTER LOADING NO ALARM HORN
LSS 314	N/A	PORTABLE	UPA 02-214 01-ID-28-001-9-D3	MUN CORR RM 05-210 LSS 26B	NOT REQUIRED	N/A	TWA	N/A	61'	PRIMARY	
LSS 315	N/A	PORTABLE	UPA 02-214 01-ID-28-001-9-C3	MUN CORR RM 05-210 LSS 26C	NOT REQUIRED	N/A	TWA	N/A	66'	PRIMARY	
LSS 316	N/A	PORTABLE	UPA 02-214 01-ID-28-001-9-B2	ECV RM 04-213 LSS 30A	NOT REQUIRED	N/A	TWA	N/A	35'	PRIMARY	
MUN 317	322	N/A	OBS COR 09-207 01-ID-28-001-8-C3	MUN COR 05-210	UPS	A/B	TWA	40	100'	N/A	SAMPLE LINE RELOCATED AS NEEDED
ECR 320	327	N/A	OBS COR 09-204 01-ID-28-001-9-B4	ECR "A" RM 03-211	UPS	A	ECL TWA	40	100'	N/A	SAMPLE LINE RELOCATED AS NEEDED
LSS 318	N/A	PORTABLE	OBS COR 09-203 01-ID-28-001-8-A4	MUN COR 05-210 34D	NOT REQUIRED	N/A	TWA	N/A	80'	PRIMARY	
AL 351	316	325	OBS COR 09-142 01-ID-28-001-2-C4	AIRLK "C" RM 06-139	UPS	C	TWA	0.2	30'	ACAMS CONFIRM	

AGENT MONITORING PLAN
GB ROCKETS, TONS, AND 105 mm PROJECTILES, AND HD MONITORING FOR LAB P&A

STA #	ACAMS TAG #	DAAMS TAG #	UNIT LOC FLOOR PLAN # DRAWING #	AREA MONITORED FLOOR PLAN #	POWER TYPE SPS/UPS	SAMPLE POINT HAZ CAT	MON-ITORING LEVEL	ALARM LEVEL	SAMPLE LINE LENGTH ±10%	DAAMS MODE	COMMENTS
LSS 307	N/A	PORTABLE	OBS COR 09-207 01-ID-28-001-9-A3	ECV RM 04-213 LSS 25B	NOT REQUIRED	N/A	TWA	N/A	30'	PRIMARY	
LSS 308	N/A	PORTABLE	OBS COR 09-207 01-ID-28-001-9-A3	ECV RM 04-213 LSS 25D	NOT REQUIRED	N/A	TWA	N/A	10'	PRIMARY	
LSS 310	N/A	PORTABLE	UPA 02-214 01-ID-28-001-9	UPA RM 02-214 LSS 44B	NOT REQUIRED	N/A	TWA	N/A	1'	PRIMARY	LSS MANIFOLD NOT USED STA. INACTIVE.
ECR 311	219	128	OBS COR 09-204 01-ID-28-001-9-B4	ECR DUCT "A" RM 03-211	UPS	A	ECL IDLH WITH LVS	CON 0.2 Z LOC 0.2 Z	70'	HIST SYST	NO ALARM HORN
ECR 312	105	120	OBS COR 09-204 01-ID-28-001-9-B4	ECR DUCT "B" RM 03-212	UPS	A	ECL IDLH WITH LVS	CON 0.2 Z LOC 0.2 Z	70'	HIST SYST	NO ALARM HORN
DUC 313	180	N/A	UPA 02-214 01-ID-28-001-9-B2	90" EXH DUCT	UPS	NA	ECL IDLH WITH LVS	CON 0.2 Z LOC 0.2 Z	200'	N/A	MONITORED TO CALCULATE FILTER LOADING NO ALARM HORN
LSS 314	N/A	PORTABLE	UPA 02-214 01-ID-28-001-9-D3	MUN CORR RM 05-210 LSS 26B	NOT REQUIRED	N/A	TWA	N/A	61'	PRIMARY	
LSS 315	N/A	PORTABLE	UPA 02-214 01-ID-28-001-9-C3	MUN CORR RM 05-210 LSS 26C	NOT REQUIRED	N/A	TWA	N/A	66'	PRIMARY	
LSS 316	N/A	PORTABLE	UPA 02-214 01-ID-28-001-9-B2	ECV RM 04-213 LSS 30A	NOT REQUIRED	N/A	TWA	N/A	35'	PRIMARY	